




Test Certificate

Part 1 Liter Vessel
Part Number 14322-01
Serial Number 98110 - 98151



		USP / Ph.Eur. Tolerances		
Dimensions	Vessel	Inside Ø a	98.5 to 101.5 mm	98 – 106 mm
		Height b	200 mm	160 – 210 mm
		With hemispherical bottom		
Material	Borosilicate glass			

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopoeia requirements (at date of approval, see below).

Aesch, 2024-04-16

Michael Schmidlin
SOTAX Quality Management
SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland